

Certificate of Analysis

Material: 87627260 **Acide chlorhydrique 0.02 mol/l (0.02 N)**

Batch : 2602596 **Expires end of** 24/08/2027

CHARACTERISTICS	SPECIFICATIONS	MESURED VALUES
Titer (20°C; real value 0.3 % accuracy)	0,01996 --> 0,02004 mol/l	0,02000 Mol/l

We Certify that this batch conforms to the specifications listed above
BDL : Below detected limit

Date

01/09/2026

Signé par:

A handwritten signature in black ink, appearing to read "Bringer", written over a horizontal line.

O. BRINGER , Responsable Qualité