

Certificate of Analysis

Material: MC3006511000 **Ammonium chlorure 0,1 N**

Batch : 2602611 **Expires end of** 25/08/2027

CHARACTERISTICS	SPECIFICATIONS	MESURED VALUES
Titer (20°C; real value 0.3 % accuracy)	0,0998-->0,1002 mol/l	0,1000 mol/l

We Certify that this batch conforms to the specifications listed above
BDL : Below detected limit

Date

01/09/2026

Signé par:

O. BRINGER , Responsable Qualité