

Certificate of Analysis

Material: MC3125561000 **Potassium permanganate 0.04 mol/l (0.2 N)**
en solution aqueuse

Batch : 2602702 **Expires end of** 03/09/2027

CHARACTERISTICS	SPECIFICATIONS	MESURED VALUES
Titer (20°C; real value 0.3 % accuracy)	0,03992 --> 0,04008 mol/l	0,04002 mol/l

We Certify that this batch conforms to the specifications listed above
BDL : Below detected limit

Date

07/09/2026

Signé par:

A handwritten signature in black ink, appearing to read 'O. Bringer', written over a horizontal line.

O. BRINGER , Responsable Qualité