

Certificate of Analysis

Material: MC3125579010 **Acide sulfurique 1,25 mol/l (2,5 N) en solution aqueuse**

Batch : 2602706 **Expires end of** 04/09/2027

CHARACTERISTICS	SPECIFICATIONS	MESURED VALUES
Titer (20°C; real value 0.3 % accuracy)	1,248 --> 1,252 mol/l	1,250 mol/l

We Certify that this batch conforms to the specifications listed above
BDL : Below detected limit

Date

07/09/2026

Signé par:

A handwritten signature in black ink, appearing to read "Bringer", written over a horizontal line.

O. BRINGER , Responsable Qualité