Certificate of Analysis

avantor™ delivered by VWR™

Material Number:	88571.290
Material description :	Nitric acid 8 mol/l (8 N)
Grade :	AVS TITRINORM®
Lot Number :	2602991
Use by* :	2027-Sep-29
Date of Quality Control :	2026-Sep-29
Date of Manufacture :	2026-Sep-29
CAS Number :	
Molecular weight:	
Molecular formula:	

Characteristics	Specifications	Measured values
Titer (20°C; real value 0.3 % accuracy)	7,99-->8,01 mol/l	7,99 Mol/l

Signature

We certify that this batch conforms to the specifications listed above.This document has been produced electronically and is valid without a signature.

Signed on behalf of VWR 2026-Sep-29

O Bringer
Quality Control Manager

Other information

*The use by date is the date to which VWR guarantees the above specification will be met for the product in an unopened container stored und recommended conditions.